Republika e Kosovës
Republika Kosova - Republic of Kosovo
Qeveria – Vlada - Government
Ministria e Shëndetësisë - Ministarstvo zdravstva - Ministry of Health
Divizioni i Prokurimit/Divizia Nabavke/Procurment Division

Prishtine datë, 01.09.2023

Zyrtari Përgjegjës i Prokurimit i Ministrisë se Shëndetësisë, në mbështetje të nenit 108/A (paragrafi 10.1) të Ligjit per Prokurimin Publik nr. 04/L-042, i ndryshuar dhe plotësuar me ligjin Nr. 04/L-237, ligjin Nr. 05/L-068 dhe ligjin Nr. 05/L-092, duke u bazuar në nenin 63 të Rregullores per Prokurim Publik të aprovuara nga Komisioni Rregullativ i Prokurimit Publik, duke vendosur sipas **kërkesës për ri-shqyrtim** të datës **25.08.2023** dhe **kërkesës për ri-shqyrtim** të datës **29.08.2023,** të parashtruar nga Operatori Ekonomik **KEIS Pharmaceutical Sh.P.K.,** lidhur me aktivitetin e prokurimit me titull: **"Furnizim me pajisje mjekësore, instrumente mjekësore për Njësitë organizative te QKUK-së" Lot 1, 3, 31, 32, 33, dhe 34,** me numër prokurimi **206-23-4215-1-1-1,** nxjerr:

V E N D I M

1. **Refuzohen** kërkesat për ri-shqyrtim e operatorit ekonomik KEIS Pharmaceutical Sh.P.K., te dt. 25.08.2023 dhe te dt. 29.08.2023, lidhur me aktivitetin e prokurimit "Furnizim me pajisje mjekësore, instrumente mjekësore për Njësitë organizative te QKUK-së" Lot 1, 3, 31, 32, 33 dhe 34, me numër prokurimi 206-23-4215-1-1-1, kundër Njoftimit mbi Vendimin e AK te datës 21.08.2023, te publikuar ne platformën elektronike;

2. **Mbeten ne fuqi** Njoftimi mbi Vendimin e AK, i datës 21.08.2023;

3. Autoriteti Kontraktues, vazhdon tutje me aktivitetin e prokurimit;

4. Ky vendim hyn në fuqi ditën e nënshkrimit dhe do të publikohet për të gjitha palët në platformën elekronike në adresën : https://e-prokurimi.rks-gov.net.

A r s y e t i m

Operatori ekonomik KEIS Pharmaceutical Sh.P.K., mr dt. 25.08.2023 dhe 29.08.2023 ka parashtruar kërkesa për ri-shqyrtim, lidhur me aktivitetin e prokurimit me titull: "Furnizim me pajisje mjekësore, instrumente mjekësore për Njësitë organizative te QKUK-së" – Lot 1, 3, 31, 32, 33 dhe 34, me numër prokurimi 206-23-4215-1-1-1. Autoriteti Kontraktues, e ka shqyrtuar kërkesën në fjalë si më poshtë:

Dispozitat e LPP-së që supozohet të jenë shkelur nga AK sipas OE KEIS Pharmaceutical Sh.P.K:

- Neni 1 Qëllimi
- Neni 7 Barazia nëTrajtim/Jo-Diskriminimi

- Neni 28 Specifikimet teknike
- Neni 59 Ekzaminimi,Vlerësimi dhe Krahasimi i Tenderëve
- Neni 60 Kriteret për Dhënien e Kontratës
- Neni 69 Aftësia Teknike dhe Profesionale
- Neni 72 Dokumentacioni dhe Informacioni Shtesë

Pretendimet Ankimore te OE KEIS Pharmaceutical Sh.P.K:

Informata te pergjithshme byrese:

Aktiviteti I prokurimit: Furnizim me pajisje mjekesore, instrumente mjekesore per njesite organizative te QKUK-se si procedure ka zgjatur shume me teper sesa koha e parashikur dhe percaktuar ne dosje te tenderit. Ky aktivitet I prokurimit eshte pezulluar disa here perkak te kerkesave te shumta per pyetje sqaruese. Ne shumicen e pergjegjieve te pyetjeve sqaruese AK eshte pergjigjur pozitivisht.

Rrethanat faktike te Llotit 1, 3, 31, 32, 33, dhe 34.

Lot#1 Arsyet e eliminimit:

Kapaciteti teknik dhe professional Kërkesa 2. Operatori ekonomik te prezantoj katalogun (broshurën) e pajisjeve te ofruara dhe te nenvizohen ne menyre te qarte produktet dhe specifikat teknike te ofruara. Nëse ofrohet broshura, duhet te specifikohet linku i veb faqes.

KEIS Lot#1:

ARSYETIM

Ne brochuren e ofruar ne dokumentacionin e aplikimit te tenderit-file I emertuar KEIS Lot#1, ART 1 Brochure, CE, MAF, ne faqet 9, 12-19 eshte I nenvizuar qarte modeli I ofertuar ndersa nen model tregohet qarte se cili model cfare specifika permban apo specifikimet teknike qe perfshine ai model, ndersa ne faqen 19 eshte edhe linku ku mund te kyqeni ne brochuren e ketij prodhuesi. Njjete jane per te gjithe artikujt dhe akcesoret tjere te ofertuar per Llotin#1 (KEIS Lot#1, ART 2 Brochure, CE, MAF dhe KEIS Lot#1, ART 3 dhe 4 Brochurat, CE, MAF. Te nenvizosh cdo fjale apo fjali mendojme se eshte vetem ceshtje ilustrative qe nuk ka ndonje ndikim ne funksionin e pajisjes si dhe nuk permban baze ligjore per tu eleminuar nga tenderi.

Ne vazhdim jane deshmite e nenvizimit te specifikimeve teknike te secilit produket te Llotit 1 dhe faqet e nenvizuara te produkteve, file I emertuar si KEIS Lot#1, ART 1 Brochure, CE, MAF,.

Lot#1, ART 1

- Intro
- The Advantages
- The Applications
- **The System**
- Technology and Details
- Service



1 Microscope

- Axio Observer 3 materials (encoded)
- Axio Observer 5 materials (encoded, partly motorized)
- Axio Observer 7 materials (motorized)

Faqe 9

Technical Specifications

Equipment	Options	Axio Observer 3 materials
Stand	Manual	●
	Motorized	
Accessory	Separable by PC	●
Display	LCD display	
	TFT display	
	Docking Station	
Interfaces	CAN	●
	PS 232	●
	USB	●
	ICAN	
	Socket for external USB(1) [illegible]	
	Trigger socket [illegible]	
CAN bus, 4 pos.		
Light Manager		●
Contrast Manager		
Circular Operation Key Unit	Light	
	Left	
Z-focus Drive	Manual [illegible]	●
	Motorized, [illegible]	
Adjustable Limit Stop for Z-focus	Manual	

Technical Specifications



Technical Specifications



Page 14

Technical Specifications



Page 15

Technical Specifications



Page 16

Profit from the optimized performance of your microscope system with services from ZEISS – now and for years to come

=> www.zeiss.com/microservice

19

Page 19



Page 35

Microscope compatibility table (—select model based on your microscope)

	Olympus		Nikon		Carl Zeiss		Leica	
MTK-1/4	O1	IX-70	N1	ECLIPSE TE2000	Z1	Axiovert 200	L1	[illegible]
	O2	IX-71	N2	ECLIPSE Ti		Axio Observer		
	O3	IX-71/83	N3	ECLIPSE Ts2R	Z2	Axio Vert A1	L2	DMR

Specifications are subject to change without notice.

NARISHIGE

Distance for your solutions

NARISHIGE SCIENTIFIC INSTRUMENT LAB.
27-8, Mizuhanasuyama 4-chome, Setagaya-ku, Tokyo 157-0062, Japan
Tel: +81 (0)3 3308 6000 Fax: +81 (0)3 3308 2000
e-mail: sales@narishige.co.jp

Dealer/Distributor

Page 36

NARISHIGE

IM-11-2



Page 40

OCTAX LASER AND IMAGING SYSTEMS

Efficient, safe and trusted for all laser applications in IVF.

Page 55

OCTAX LASERSHOT
Collimator
20x ~ 40x Objective
Control Unit
Laser Power Supply

Page 66

NARISHIGE

IM-11-2
Pneumatic Microinjector

Smooth and comfortable pneumatic control for both injecting and holding sides

The IM-11-2 is a pneumatically operated microinjector that allows suction and injection of a microscopic with high precision and operability comparable to oil hydraulic injectors. Equipped with a coarse-fine coaxial control knob, the microinjector enables both coarse and fine movement operation. The IM-11-2 has substantially improved on the amount of pressure change per knob rotation, thereby enabling finer and more powerful control.



Page 69



Page 73

MiA AiR

AIR PURIFIER



Specifications	Plus	Compact
Max. Clean Air Delivery Rate*	360 CADR	250 CADR
Effective Room Coverage (m²)	60	35
Mobile App/Wi-Fi	Yes	Yes
Control Panel	Capacitive Touch	Manual Touch
Sensors	PM, VOC, Light, etc.	Yes
Fan Speed	6	3
Automatic Modes	6	3
Filtration		
Filter Thickness (H13)	440 mm	290 mm
Activated Carbon (kg)	2	0.8
Unique Filter Area (m²)	8	3
HEPA 13	Yes	Yes
HEPA 13	Yes	Yes
HEPA 13	Yes	Yes
Other Features/Indicator		
Mobile App	Yes	No
Child Light on Control	Yes	Yes
Remote Control	Mobile App	
Real Time Indoor Air Monitoring	Mode + On Display	No
Colors	White, Anthracite, Red	White, Anthracite, Red
Noise Level (Auto Mode)	56 dB(A)	50 dB(A)
Size (W/H/D) (mm/inch)	440 x 580 x 300 17.32 x 22.83 x 11.81	290 x 580 x 290 11.42 x 22.83 x 11.42
Weight (kg)	17	9
UV Light	Optional	No
Max. Power Consumption (W)	80 + 10 (UV)	50

www.mia-air.com

Page 81



Exo Medical AVT-1® is an anti-vibration table, intended to be used with high-magnification microscopes. It isolates the microscope from vibrations generated in the room where the table is placed and from vibration generated above or below any part of the table top, that is not part of the AVT platform.

The table consists of a sturdy frame made from high-quality stainless steel, which makes it easy to clean and maintain. The tabletop is made of stainless steel with glass on both sides.

Page 83

Ne vazhdim jane deshmite e nenvizimit te specifikimeve teknike te secilit produket te Llotit 1 dhe faqet e nenvizuara te produkteve. file I emertuar si KEIS Lot#1, ART 2 Brochure, CE, MAF..



Faqe 9

Technical Specifications

Faqe 12

Technical Specifications



Technical Specifications



Technical Specifications



Technical Specifications



Technical Specifications

Page 17



Profit from the optimized performance of your system with services from ZEISS – how brief for

>> www.zeiss.com/microservice

Page 18



Page 35

Specification

Description / Model	Fully motorized Manual Stage with Joystick Ultra-precision Microscope Stand MIX-1xx

Dimensions / Weight

Control Unit	W : 45mm	D : 330±5mm	H : 131 ~ 170mm	Weight : 1,085g
Drive Unit	W : 185 ~ 435mm	D : 102±5mm	H : 100 ~ 158mm	Weight : 8.18g

Drive method

	Address	Range	Operation		Performance	Angle	Error Free compensation range
X	01	Basic	Axial Hypoid Gear	Ball Screw	250 μm		
Y	01	Basic			250 μm		
Z	01	Option	Worm		250 μm		
Θ (Objective Rotation)	01	Option	Worm		1,000 μm	15 ~ 45°	
	Mechanical	Option	Lever control (with Boost)				

Microscope combinability table (○=selected model based on your requirements)

	Options		Models		Carl Zeiss		Leica	
MIX-1xx	01	IX-70	91	ECLIPSE TE2000	21	1 Axiovert 200 2 Axio Observer	L1	DMIRB/E/M/RE/IRM/IRE/B
	02	IX-71		ECLIPSE Ti			L2	DMI6
	03	IX-73/83			22	Axio Vert.A1		

Specifications are subject to change without notice.

MARSHGE

MARSHGE CO., LTD.
21-9, Musashiseki-yama 4-chome, Setagaya-ku, Tokyo 157-0067, Japan
Tel: +81 (0)3 3300 5331 Fax: +81 (0)3 3300 5700
e-mail: sales@marshge.co.jp
MARSHGE INTERNATIONAL USA, INC.
1710 Hempstead Turnpike, East Meadow, NY 11554, U.S.A.
Tel: +1 516 794 3500 Fax: +1 516 794 0555 Toll Free: 888 445 7814
e-mail: info@marshge-usa.com
MARSHGE INTERNATIONAL LTD.
Unit T, Millers Business Park, Yellow Way, London NW10 6SP, U.K.
Tel: +44 (0) 20 8830 1666 Fax: +44 (0) 20 8 391 9555
e-mail: sales@marshge.co.uk
Web site: http://marshge-group.com

Distributor

Page 36

Ne vazhdim jane deshmite e nenvizimit te specifikimeve teknike te secilit produket te Llotit 1 dhe faqet e nenvizuara je produkteve, file I emertuar si KEIS Lot#1, ART 3&4 Brochure, CE, MAF..



Art 3, faqe 1



Art 3, faqe 34



Art 3, faqe 36

ZEISS Stemi 508

Your Apochromatic Stereo Microscope with 8:1 Zoom for Excellent Image Contrast and Color Accuracy

Art 4, Page 42

Tailored Precisely to Your Applications



Art 4, Page 54



Art 4, faqe 55



Art 4, faqe 71

Lot#1.

Gjitheashtu sa I perket po kesaj pike si arsyetim tjeter se PSE nuk eshte dashur te eleminohemi per arsyen e dhene eshte edhe kolona e fundit e paraqitur ne dosje te tenderit: Dëshmia e kërkuar dokumentare: ku kerkohet **Dëshmia 1. Katalogu (broshura) e pajisjeve te ofruara me selektim te qarte te produkteve dhe specifikave teknike te ofruara.** Ne si kompani e e kemi plotesuar ne teresi kete kriter.

KEIS Lot#32:

Art:5 ARSYETIM

Artikulli I ofertuar nga prodhuesi ESO Medical, eshte pajisja nje nga pajisjet me te mira ne tregun boteror te dedikuar per perdorim ne IVF. Vet prodhuesi ESCO Medical eshte I certifikar dhe operon vetem me produket te dedikuara per perdorim ne fushen e IVF prandaj asnje nga arsyetimet per eleminim mendojme se nuk qendron per kete pozicion prandaj kerkojme nje rishqyrtim te dokumentacionit dhe certifikatave te cileses sidomos CE dhe Deklaratave te konformitetit ne te cilat percaktohet qarte se pajisja e ofertuar conform cilave standarve operon: psh CE dhe FDA jane certifikata te njejta te cilesise vetem se njera eshte e dedikuar per pajisjet per tregun European ndersa tjetra per tregun amerikan.

Ju lutem referojuni file te emertuar: KEIS Lot#32, ART 5,6,7 Brochure, CE, MAF, faqe 6

General Specifications

Esco Multi-Zone ART Workstation (MAW)

Model		MAW-40	MAW-60 , MOBO	MAW-60 , DUAL	MAW-60 , MP
Nominal Size		1.2 meter (4')	1.8 meter (6')		
Internal Work area dimensions (width x depth x height)*		1200 x 500 x 710 mm (49.6" x 19.7" x 28.0")	1800 x 500 x 710 mm (71.6" x 19.7" x 28.0")		1800 x 650 x 750 mm (74.4" x 19.3" x 30.7")
External dimensions without support stand (width x depth x height)		1340 x 640 x 1300 mm (52.8" x 25.2" x 51.2")	1950 x 640 x 1300 mm (76.8" x 25.2" x 51.2")		1950 x 640 x 1300 mm (76.8" x 25.4" x 53.5")
External dimensions with "B" type support stand (width x depth x height)		1340 x 640 x 2100 mm (52.8" x 25.2" x 85.0")	1950 x 640 x 2100 mm (76.8" x 25.2" x 85.0")		1950 x 640 x 2100 mm (76.8" x 25.4" x 87.4")
Cabinet Construction	Main Body	1.2 mm (0.05") 16 gauge electro-galvanized steel with white oven-baked epoxy polyester powder coated finish			
	Work Zone	3.2 mm (0.05") 16 gauge stainless steel, grade 304			
	Side Walls	Tempered Laminated Glass (Optional: Polycarbonate sash for MAW with built-in UV light)			
	Sash				
Laminar air velocity		Average of 0.31m/s or 41 fpm (± 20%)			
Filter efficiency		>99.999% for particle size between 0.1 to 0.3 microns per IEEE-PC2001.3 / 1994 per EN 1822			
Pre-filter		Disposable and non-washable polyester fibers with 85% arrestance (ASD) rated			
Noise level (per NSF 49)**		47 dBa	52 dBa		
Set of (9+1) heating zones		1 set	1 set	2 sets	1 set
Surveillance System with data logger		1 set	1 set	2 sets	1 set
Humidification System***		1 set	1 set	2 sets	1 set
Microscope		Required Not included (see microscope ordering information)			
		Position for 1 microscope		Position for 2 microscopes	Position for 1 microscope and 1 inverted microscope
Support stand		Required Not included (see support stand ordering information)			
Shipping dimensions (width x depth x height)		1450 x 850 x 1520 mm (57.1" x 33.5" x 75.6")	2050 x 850 x 1800 mm (80.8" x 33.5" x 74.4")		
Shipping weight		280 kg	372 kg		

Faqe 6

Art: 6 dhe 7. ARSYETIM

Qe nga momenti I shpjalles se aktivitetit te prokurimit AK ligjerisht eshte I obliguar qe te filloje dhe vazhdoje komunikimin e vazhdueshem me te gjithe OE e interesuar per te marre pjese ne ate aktivitet te prokurimit (kerkesa per sqarim te DT, rishqyrtim etj etj). Sa I perket katalogjeve te ofruara ne dokumentacion e aplikimit AK eshte dashur qe ten a kerkoje sqarime shtese ose forma te tjera paraqitjes se katalogjeve psh hard copy dhe jo te na eleminoj per mos lexim te tyre. Ne kemi qene dhe jemi te gatshem ti dorezojme katalogjet prane AK ne cilen do forme te mundshme te cilat jane te kupruesheme dhe lexueshem per AK. Gjitheashtu ne katalogjet e ofruara ka qene edhe kontakti I prodhuesit (perfshire edhe website) ne te cilin AK ka mundur ta shkakrtoje edhe vet vet katalogun e ofruar. Artikujt e ofertuar jane te deshmuar se jane ekskluzivisht te dedikuar per perdorim ne fushem e IVF dhe kryejn te gjitha funksionet e kerkuara pa asnje kufizim sikur te pajisjet e te gjithe prodhuesve te tjere boteror.

Ndersa sa I perket kriterit te eleminimit per mungese te deshmise se disponueshmerine e pjeseve reserve per se paku 7 vitet e ardhshme, kjo nuk qendron sepse prodhuesi e ka ofruar kete garancion te disponueshmerise se pjeseve reserve per 7 vitet e ardhshme ne autorizimin e tij.

Ju lutem referojuni file te ennertuar: KEIS Lot#32. ART 5,6,7 Brochure, CE. MAF, foto'a ne vazhdim faqe 47:

Authorization Statement

Date: 19.08.2021

To: MINISTRY OF HEALTH

Ne perkohesisht	2021-21-4215-1-15
Autorizohet	

Subj: "Supply of medical equipment, medical instruments for the organizational needs of the OEM"

WHEREAS

We, MIKROTEST LABORATUVAR CİHAZLARI MAKİNE İMALAT SAN. TİC. LTD. ŞTİ., with an official manufacturer of Laboratory tools and equipment, laboratory equipment and accessories of same duties, having facilities at [Oz Ankara Sanayi Sitesi 1478 (OSB) Sok. No:37 Lodumlu/Gölbaşı-ANKARA], do hereby authorize *KEIS Pharmaceutical Sh.p.k with Unique Identification Number J11811902I* to submit a quotation the purpose of which is to provide the following Goods, manufactured by us and to subsequently negotiate and sign the Contract.

Manufacturer's warranty for availability of spare parts at least for 7 years from the date of installation

Dated on 19/08/2021

MIKROTEST LABORATUVAR CİHAZLARI MAKİNE İMALAT SANAYİ TİCARET LİMİTED ŞİRKETİ

KEIS Lot#33:

ARSYETIM: Ne momentin e shpalljes se ketij aktiviteti te prokurimit te gjithe artikujt e llotit 1, 32, 33 dhe 34 kane qene te perfshira vetem ne nje lot (LOT#1) dhe vlera e buxhetit e dedikuar per kete llote ka qene: 826.900,00 euro.

Pas kerkeses se OE ekonomik per ndarje te Llotit 1 ne disa llote te tjera, bashke me ndarjen e ketij llоti ne 4 llote eshte bere edhe ndarja e buxhetit per secilin llot per te cilin mendojme se ndarja e buxhetit per keto 4 llote eshte bere ne menyre jo proporcionale qe nenkupton per ndonje llot eshte ndare me shume buxhet sesa eshte dashur ndersa per ndonje llot me pak sesa eshte dashur.

Per kete ndarje jo propocionale te buxhetit ka pasur edhe kerkese per rishqyrtim te buxhetit.

Edhe perkunder ndarjes jo propercionale te buxhetit kompania KEIS Pharmaceutical per te 4 llotet e IVF ka aplikuar ne vlere totale 808.990,80 qe nenkupton jemi mbremda buxhetit te parapre ne fillim. Prandaj mendojme se edhe ky arsyetim per eleminim nuk qendron sepse nuk eshte ne pyetje tejkalimin enorm I buxhetit te parapare por ndarja joproporcionale e buxhetit per llote.

1. Ju lutem referojuni sqarimeve shtese te AK te cilat jane te disponueshme ne faqen e-prokurimi dhe ate si vijon:

   Linku:

   https://e-prokurimi.rks-gov.net/HOME/ClanakItemNew.aspx?id=337

   Dokumentet-emertimi dhe datat:

   VENDIM PER RISHQYRTIM TE KERKESES_Access Solution shpk te dates 15.06.2023.pdf

Republika e Kosovës
Republika Kosova - Republic of Kosovo
Qeveria – Vlada - Government
Ministria e Shëndetësisë - Ministarstvo zdravstva - Ministry of Health
Departamenti Prokurimit/Divizia Ndërsektoriale/Procurement Division

Prishtinë më/dne, 15.06.2023

Zyrtari Përgjegjës i Prokurimit i Ministrisë së Shëndetësisë, në mbështetje të nenit 108/A paragrafi 108.1 të Ligjit për Prokurimin Publik nr. 04/L-042, i ndryshuar dhe plotësuar me ligjin Nr. 04/L-237, ligjet Nr. 05/L-068 dhe ligjin Nr. 05/L-092, dhe të Udhëzimit të nenit 65 të Rregullores për Prokurim Publik të aprovuara nga Komisioni Rregullativ i Prokurimit Publik, duke vendosur sipas **Kërkesës për të shqyrtuar** të datës 12.06.2023, të parashtruar nga Operatori Ekonomik **Access Solution shpk**, sihtuar me ankimin e prokurimit me titull "**Furnizim me pajisje mjekësore, instrumente mjekësore për Njësitë organizative të QKUK-së**" **LOT 1, 32, 33, 34**, me numër prokurimi **208-23-4218-1-1-1** nxjerr

V E N D I M

1. **Pranohet** kërkesa për rishqyrtim e operatorit ekonomik Access Solution shpk e të 12.06.2023 sihtuar me ankimin e prokurimit me titull "Furnizim me pajisje mjekësore, instrumente mjekësore për Njësitë organizative të QKUK-së" Lot 1, 32, 33, 34, me numër prokurimi 208-23-4218-1-1-1, kundër Njoftimit për Kontratë të datës 10.04.2023 të publikuar në platformën elektronike.
2. Do të rishikohet, Specifikimi Teknik dhe buxheti për Lot 1, 32, 33, 34
3. Autoriteti Kontraktues, e shfryt afatin për plotësim për shkaqeve e publikimit.
4. Ky vendim hyn në fuqi ditën e marrjes së tij dhe do të publikohet për të gjitha palët në platformën elektronike të qeverisë: https://e-prokurimi.rks-gov.net

A r s y e t i m

Operatori Ekonomik Access Solution shpk me datë 12.06.2023 ka parashtruar kërkesën për rishqyrtim ndaj dhënies së prokurimi me titull "Furnizim me pajisje mjekësore, instrumente mjekësore për Njësitë organizative të QKUK-së" – Lot 1, 32, 33, 34, me numër prokurimi 208-23-4218-1-1-1. Autoriteti Kontraktues e ka shqyrtuar kërkesën në fjalë si në vijim:

Dispozitat e LPP-së në mbështetje të të cilës Autoriteti Kontraktues ka vendosur sipas OE Access Solution shpk:

Pretendimet ankimore:

Pas së shkruajturës dhe kërkesës për të mbrojtur interesin e Juridik të Lot 1 - Furnizim me pajisje mjekësore për Familjen të Aziksur Mjekësor IVF në Kl. e Gynekologjisë me Onkologji. Juridik i paraqitshëm për këto llot pasi ndahen në 4 llote se pava ka qenë 826.900 euro.

KEIS Lot#34:

ARSYETIM: Te gjitha arsyetimet e dhena per eleminim per kete lot nuk qendrojne per arsyet vijuese:

2. Per lotet e IVF disa here ka pasur kerkesa per sqarim per specifikat teknike dhe pergjegjiet e AK per keto kerkesa kane qene pozitive: LEJOHET TE APLIKOHET ME PRODUKTE ME SPECIFIKA TEKNIKE TE NGJASHME DHE QE E KRYEJNE TE NJEJTIN FUNKSION. Ju lutem referojuni sqarimeve shtese te AK te cilat jane te disponueshme ne faqen e-prokurimi dhe ate si vijon:

   Linku:

   http://e-prokurimi.rks-gov.net/HOME/ClanakItemNew.aspx?id=327

   Dokumentet-emertimi dhe datat:

   Sqarime shtese per OE te dates 26.05.2023.pdf

   Sqarime shtese per OE te dates 29.05.2023.pdf

   Sqarime shtese per OE te dates 07.06.2023.pdf

   Sqarime shtese per OE te dates 12.06.2023.pdf

   VENDIM PER RISHQYRTIM TE KERKESES_INOVIX_ te dates 27.06.2023.pdf

   Sqarime shtesë per Lot 1,33,34_te dates 27.06.2023.pdf.

   Foto e vendimit ne vazhdim te Sqarime shtesë per Lot 1,33,34_te dates 27.06.2023.pdf

Gjuha e softuerit te menaxhimit te shoqeruar duhet te jete ne Anglisht, por duhet te mundesohet edhe te perdoruesit te kete mundesi te ndryshojne gjuhen e softuerit ne gjuhen e tyre si psh gjuha ne anglisht kerkuesi ndryshues

Softueri e shoqeruar duhet te jete ne perputhje me standardet ISO si ISO27001. Softueri duhet te mbeshtese autentifikimin e bazuar ne platforme, autentifikimin me 2 faktore dhe autentifikimin me SMS, gjithashtu te gjitha te dhenat e informacionit personal duhet te ruhen plotesisht te koduara

Softueri i shoqeruar i menaxhimit duhet te mbeshtese versionet me te fundit te DB dhe SQL DB dhe SQL duhet te jene pjese kontratave te softuerit te menaxhimit te shoqeruar. Licencat shtese mund te menaxhohen dhe mentenohen me perputhje me ASHE, per nderr ASHE nuk i ka ose nuk i mbulon te gjitha licencat e kerkuara per arsye te regullorave te tyre ose mungeses se licencave, shtohen te gjitha licencat e kerkuara duhet te mbulohen nga furnizuesi.

Dokumentacioni i perngjitshem i funksionimit dhe perdorimit te sistemit duhet te dorezohet dhe te ruhet ne mbyrje te sigurt ne format elektronik pas instalimit te sistemit

Duhet te jete i mjaft softuer dublu paralleve ne sistemin (Main Program, ID Witness dhe Tank Kontrolli dhe duhet te jete i sinkronizuar me te njejtin server pa nevoje per server shtese.

Per Lot 1, 33, 34, ne rast te kerkese te ndonjeri te nenjesi te nese mendon pamjaftueshem, per vleresim (pas aksesimit) mbarsa per specifikat e kerkuara, prezantimin te njejta dhe me te avancuara me qellim te konkurrencis sa me te gjere ne tenderin

Me respekt,

Bardyl Mamudhi

Zyrtari përgjegjës i Prokurimit

3. Arsyetimet tuaja per nxos permbushje te disa specifikimeve teknike kane te bejne vetem me aspektin ilustrativ dhe nuk kane asnje efekt ne funksionin e te gjithe artikujve te kerkuara sepse per cdo paqartesi kemi mundur te na kontaktoni per sqarime.

ARSYETIM I PERGJITHSHEM:

Kerkojme nga komisioni vleresues qe te rishqyrtoje edhe njehere te gjithe dokumentacionin e ofruar nga kompania KEIS dhe te beje edhe njehere rivlersimin e ofertave tona sepse mendojme se I plotesojme te gjitha kriteret e kerkuara dhe te na shpallin fitues ekskluzivisht per llotete 1, 33, 33 dhe 34.

Gjitheashtu plotesimin e specifikimeve teknike funksionale e kemi konfirmuar edhe me plotesimin dhe nenshkrimin e aneksit 1 te DT. Kerkojme nga komisioni vleresuar te na jap se paku nje deshmi per ndonje nga artikujt e oferuar se ky artikull nuk eshte funksional per tu perdorur ne fushen e IVF sepse mendojme se te gjitha pajisjet e ofertuara jane te kualitetit te larte dhe te deshmuara per aplikim ne fushen e IVF.

Gjitheashtu edhe per Lot#3 mendojme se artikujt tone te ofertuar jane te kualitetit te larte dhe I plotesojne te gjitha specifikimet teknike per te qene funksional per qellimin e kerkuar.

Rrethanat Faktike të Lot#31

ARSYETIM

Në letrën standarde për tenderuesin e eliminuar në lotin 31 cekët që është oferta e kompanisë tone është e papërgjegjëshme sepse është kërkuar "Most perform single and/or double access Apharesis and Immunoadsorption - e ne kemi ofertuar me "Emergency Single needle option"

Konsiderojmë se aryetimi i juaj është i paqëndrushëm dhe në kundërshtim me dokumentet e prezantuara sepse Deklarimi që "Pajisje e ofruar dhe nuk është Pajisja që ne kemi kërkuar" është në kundërshtim me ligjin e prokurimit pasi që pajisja që kërkohet është e definuar me specifikacion dhe me qëllimin e përdorimit andaj pretendimi i juaj për brend të caktuar është i papranushëm dhe në kundërshtim me ligjin e Prokurimit. Rrjedhimisht për Most perform single and or double Apharesis and Immunoadsorption, e keni pas përshkrimin në broshurën fq 3 dhe 6. ku specifikohet konkretisht "

Ju lutem referojuni file te emertuar: KEIS Lot#31, Brochure, , faqe 3 dhe 6

1. **Operational requirements**

Specification Required:

2. Operational requirements				
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	2.2	Must perform single and/or double access Aphaeresis and Immunoadsorption.	YES Brochure pg. 3	



	Description	DN/SN	Disposable
Depletion	Depletion of platelets or platelets and lymphocytes	DN	C4L
TPE	Therapeutic Plasma exchange	DN/SN*	PL1
Adsorption	Plasma treatment with connected columns or filters	DN	PIR
RBC	Replacement or depletion of red blood cells	DN	PL1

Njëkohësishtë është cekur në letrën standarde për tenderuesin e eliminuar që është kërkuar edhe Convient Servo-controlled/CVT- ku e keni cekur që pajisja e ofruar nga ana jonë nuk e posedon këtë kërkesë. Rrjedhimisht konsiderojmë se aryetimi i juaj është i paqëndrushëm dhe në kundërshtim me dokumentet e prezantuara sepse e keni pas përshkërimin e poshtëshënuar:

2. Electricity supply

Specification Required:

6. Electricity supply				
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	6.2	Convenient servo-controlled stabilizer/CVT	YES	

Është cekur në letrën standarde për tenderuesin e eliminuar që është kërkuar edhe "List of equipment avaliable to provide calibration and routine preventive maintance support, according to the manufacture's documentation in the service/ technical manual- dhe ne si kompani nuk e kemi ofruar,

Konsiderojmë se aryetimi i juaj është i paqëndrushëm dhe në kundërshtim me dokumentet e prezantuara sepse të poshtëshënuar e keni të bashkangjitur dokumentin që e posedoni në dosjen e tenderit

Specifikimi 6.2 mbulohet nga specifikimi i UPS-së

LOT 31

Supply of equipment for the development of the apheresis process for the Nephrology Clinic

Quantity - 1 pc.

Technical specifications for therapeutic plasmapheresis and immunoadsorption

Code: UNSPSC:

Code ECRI:

1. Function description				
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**

	1.1	Therapeutic plasmapheresis is a medical treatment procedure in which the patient's blood passes through a machine - semipermeable membrane - Apheresis that separates a special component (immunoglobulin, antibodies) and returns the rest to circulation. This intervention results in a filtered plasma product that can be used to treat numerous diseases.	Immunoadsorption is used	
	1.2	Immunoadsorption is an alternative method of apheresis which selectively removes immunoglobulins while preserving other plasma proteins.	YES Brochure pg. 6	
	1.3	These interventions result in a filtered plasma product that can be used to treat many diseases.	YES Brochure pg. 8	

	2. Operational requirements			
	Sl	**Name**	**Technical specifications provided by the bidder**	**Deviation in specifications**
	2.1	Fully automatic, microprocessor controlled with touch screen access operator control panel.	YES Brochure pg. 3	
	2.2	Must perform single and/or double access Aphaeresis and Immunoadsorption.	YES Brochure pg. 3	

	3. Technical specifications			
	Sl	**Name**	**Technical specifications provided by the bidder**	**Deviation in specifications**
		Minimum requirements for new apheresis devices and, in a general form, the mechanism for evaluating the acceptability of the device hardware, software and associated apheresis sets.		
	3.1	Minimal Requirements		
	3.11	General. Equipment must be CE/UKCA/UKNI marked Manufacturers must adhere to Good Automated Manufacturing Practice (GAMP)	YES	
	3.12	**The hardware components of the device** The device must contain the following: - manual cancellation system	YES, Brochure pages 9 – 15.	

		- blood flow monitor - in-line air detector - integral blood filter - indicator of the flow of anticoagulants - the default collection volume device - visual sound alarm for the completion of the procedure - automatic standby mode for power failure - enabling self-checking to include all critical safety and operational procedures.		
	3.13	**Software components of the device** The software must provide the following parameters: - for the accepted algorithm of total blood volume calculation - for the accepted algorithm for calculating the rate of reinfusion of plasma, albumin, anticoagulant, citrate - for reinfusion of fixed upper limit anticoagulants/citrates - for the total collection volume of the programmable upper limit - should not exceed predetermined fluid reinfusion limits (eg anticoagulant, citrate, saline) - for the alarm and prevention of the use of the incorrect (incongruent) group for the programmed procedure - prevents the procedure when the predicted parameters after collection fall outside the programmable safety limits.	YES Brochure pages 16 – 18.	
	3.14	**Apheresis sets (kits).** Apheresis sets must have: - a closed system - a visual system to minimize the risk of liquid line transposition - a microbial filter in the "spiked" lines (unless the set will only be used to collect plasma for fractionation) - a diversion line and a sampling bag - a means of preventing incorrect connections to the IV fluid (eg saline) and anticoagulant kit.	YES. Technical Information Sheet 9400411 – PIR	

		- Care should be taken to include a pouch in the final package to facilitate testing for bacterial contamination		

4. System configuration Accessories, spares and consumables			
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	4.1	The system as specified-	YES	
	4.2	30-thirty disposable plasmapheresis/immunoadsorption kits and sets should be provided with the equipment	YES	
	4.3	All consumables necessary for the installation and standardization of the system are provided free of charge.	YES	

5. Environmental factors			
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	5.1	The unit must be capable of being continuously stored at an environment temperature of 0-50 degrees C and a relative humidity of 15-90%	YES Brochure page 19.	
	5.2	The unit must be able to operate continuously at temperature of 10-40 degrees C and a relative humidity of 15-90%.	YES Brochure page 19.	

6. Electricity supply			
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	6.1	The input power must be 220-240 VAC, 50 Hz for connecting to a standard plug applicable in Kosovo	YES	
	6.2	Convenient servo-controlled stabilizer/CVT	YES	
	6.3	UPS suitable for the relevant equipment with maintenance free battery for minimum backup of one hour shall be supplied with the unit.	YES	

7. Standards and Safety				
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	7.1	The product Must be FDA, CE, UL or BIS approved	YES	
	7.2	3 year comprehensive warranty and 5 year AMC after warranty	YES	
	7.3	Electrical safety complies with IEC-60601 / IS-13450 electrical safety standards	YES	
	7.4	The manufacturer must be ISO certified for quality standards	YES	
	7.5	Training in the center where the device is used for 5 doctors and 5 nurses who will start applying the procedure and comprehensive training for the staff in the center up to the recognition and functionalization of the system	YES	
	7.6	Guarantee for servicing the device and long-term supply of consumables for plasmapheresis and immunoadsorption	YES	

8. Documentation				
	SI	Name	**Technical specifications provided by the bidder**	**Deviation in specifications**
	8.1	Calibration and inspection certificate from the factory	YES	
	8.2	List of equipment available to provide calibration and routine preventive maintenance support, according to the manufacturer's documentation in the service/technical manual	YES	
	8.3	List of important spare parts and accessories with their part number and cost	YES	
	8.4	Daily, weekly, monthly and quarterly maintenance checklist instructions log book. Job description	YES	

Specifikimi 6.2 mbulohet nga specifikimi i UPS-së

Special Conditions

– Minimum 3 (three) year warranty for the equipment offered which has a value of ≥ one thousand euros per unit (during the warranty period all replaceable parts must be replaced according to the manufacturer's recommendations in certain periods of time including updates), yes during this period, the supply contractor is obliged to avoid any defect occurring in the equipment, even with the supply of new parts. As for the equipment offered with a value < one thousand euros per unit, the validity of the Warranty must be at least 1 (one) year.

– Installation and functionalization of equipment, especially those that exceed values ≥ one thousand euros per unit, training of personnel for use and basic servicing by the winning EO - to be proven with a declaration in the bid (an original declaration from the bidder).

– Upon delivery of the equipment, the winning EO must deposit the basic user and servicing manual, while for the equipment in which the software program is integrated, the Software (CD-software), the License (CD of Licenses) must also be deposited as and Backup-CD after installation of each device with its own serial number - evidenced by a declaration in the offer (an original Declaration from the bidder).

– Delivery time 45 days.

– Additional criteria for fixed devices which are provided for static assembly:

- The contractor must supply, assemble and install the equipment, including in the bid the installation works and all necessary accessories, adapters, equipment and materials.

- The supplier must ensure that all eventual damages will be remedied, be it damages to the floor, walls, electrical circuits as well as all other damages that may occur during the activity of supply, assembly and installation of these devices in the operating rooms.

- The reconstruction of the damaged area will have the same quality and finish as the pre-existing one.

ARSYETIM I PERGJITHSHEM:

Kerkojme nga komisioni vleresues qe te rishqyrtoje edhe njehere te gjithe dokumentacionin e ofruar nga kompania KEIS dhe te beje edhe njehere rivlersimin e ofertave tona sepse mendojme se i plotesojme te gjitha kriteret e kerkuara dhe te na shpallin fitues eksluzivisht per llotete 1,31, 32, 33 dhe 34.

Demet materiale

Shkelja e supozuar biznesin tonë përveç dëmit material fitimi nga kjo kontrate, dëm edhe më i madhë është demotivimi se në këtë vend punohet drejt.

Pergjigjia e Autoritetit Kontraktues ne lidhje me pretendimin Ankimore te OE ankues:

I konsiderojmë te pa bazuara pretendimet tuaja ankimore ne lidhje me ofertën tuaj për Lot 1, 3, 31, 32, 33 dhe 34, duke marre parasysh se gjate vlerësimit te aktivitetit, komisioni i ka verifikuar produktet e ofruara nga ana juaj dhe i ka potencuar mangësitë ne specifika. Produktet e ofruara nga ana juaj nuk i plotesojne specifikat teknike te kerkuara sipas DT-se, ju lutem refrohuni mangesive si me poshte:

Përgjigjeje për Lot 1:

Katalogu i ofruar për Artikullin 1 te Lot-it 1, përfshinë disa modele e qe është e pamundur për AK te identifikoje se çka është përfshirë ne konfigurimin e pajisjes e çka jo. P.SH.



Pra siç shihet me larte ne katalogun e ofruar ju keni nënvizuar vetëm modelin e Microscopit, mirëpo nuk i keni nënvizuar asnjë nga modelet e **Objectives, illumination, Cameras, Software and Accessories**, gjë qe për ne është e pa kjarte dhe krijon mundësi te dështimit ne realizim te kontratës.

Ne baze te këtij katalogu shihet qarte se **Objectives, illumination, Cameras, Software and Accessories** janë **me disa opsione**, dhe mos nënvizimi i këtyre opsioneve lenë për tu kuptuar se produkti mikroskop mund te liferohet pa këto pjese apo me opcionin me te dobët dhe se AK nuk ka te drejtën te kërkoj tjetrin me te mire ne momentin e pranimit.

Gjithashtu pas rishikimit te dokumentacionit është vërejtur se mungojnë specifikat te Mikromanipulatori, si ne vijim:

· **It should be controlled on x-y-z axes with the help of joystick.** (ju keni ofruar micromanipulator i cili kontrollohet nga joystic vetem ne aksin X dhe Y, ndërsa ne Z nga Knob)

Drive method							
	Method	Range	Operation		Full rotation	Angle	T-axis: Four dimensional image
X	Oil	10mm	X and Y joystick control	knob	250 µm	–	
Y	Oil	10mm		knob	250 µm		
Z	Oil	10mm	knob		250 µm		
T (Pipette holder)	Oil	10mm	knob		1,000 µm	15~40°	
	Mechanical	50mm	Lever control (up⇔Down)		–		

- In sensitive control, the motion should be done with a precision of 2 µm. (ju keni ofruar mikriomanipulator me saktesi minimale prej 5 µm, me devijim mbi 100%)

Movement range	X/Y/Z-axis	Oil hydraulic control 10mm Full rotation of knob 250µm Minimum graduation 5µm

- Motor controlled micromanipulator should be used to bring the injector into the field of view of the microscope (mungon specifika)
- The high-speed button with the manipulator should be able to move at 1.4 mm/s. (mungon specifika)
- The speed adjustment button with the manipulator speed should be adjusted sleeplessly between 0.1mm/s to 0.7mm/s. (mungon specifika)
- Pulling and pushing should be possible with the injector unit. (mungon specifika)
- Injector unit should have at least 1060 µl oil holding capacity. (mungon specifika)
- One full movement of the handle should be 500 µm and be equivalent to 10 µl oil movement. (mungon specifika)

Tek broshura e sistemit Laserik (**Artikulli 1, Laser System**) nuk është nënvizuar asnjë fjali.

Tek Artikulli 3

Ashtu siç e ka cekur OE ne ankesën e tije tek artikulli 3, është nënvizuar vetëm modeli i artikullit (ne faqen a pare te broshurës) por nuk është nënvizuar asnjë nga modelet specifike, pra ne faqen e pare shënon Primostar 3, siç shihet me poshtë janë 7 modele specifike te Primostar 3:

Tailored Precisely to Your Applications

- In Brief
- The Advantages
- The Applications
- The System
- Technology and Details
- Service

Order Number	Primostar 3: [illegible]	Primostar 3: [illegible]	Primostar 3: [illegible]	Primostar 3: [illegible]	Primostar 3: [illegible]	Primostar 3: [illegible]	Primostar 3: [illegible]	Primostar 3: [illegible]
Viewing angle	25°	25°	25°	25°	25°	25°	30°	35°
Stage drive right	x	x	x	x	x	x	x	x
FOV 20 mm	x	x	x	x	x			
FOV 22 mm						x	x	x
Fluorkühler	x	x	x	x	x			
Fullkühler						x	x	x
HAL						x	x	x
LED	x	x	x	x	x	x	x	x
Pointer	x	x	x					
Phototube			x			x	x	x
Stimes noseplece	x	x	x	x	x			
Stimes noseplece						x	x	x
Objectives D=0					4x, 10x, 40x, 100x Oil			
Objectives w/ 0.17	4x, 10x, 40x, 100x Oil	4x, 10x, 40x	4x, 10x, 40x	4x, 10x, 40x		4x, 10x, 40x	4x, 10x, 40x, 2x2	4x, 10x, 40x, 2x2
Abbe condenser	x	x	x	x	x	x	x	
Turret condenser								x
Light manager						x	x	x
Eco mode						x	x	x

Artikulli 4:
Ju keni nënvizuar modelin por jo edhe zgjedhjen e opsioneve te cilat konfigurohet Mikroskopi:



Katalogu i ofruar nga ana juaj përfshin disa modele te Microscopave, dhe mos nënvizimi i specifikave teknike te Microskopit, mos nënvizime i tyre e pamundëson një vlerësim te drejt.

<u>Për Lot 3:</u> Nuk keni ofruar asnjë pretendim ankimor.

<u>Për LOT 31:</u> Pretendimet tuaja ankimore i konsiderojmë te pa bazuara, meqenëse pajisja e ofruar nga ana juaj eshte tjetër pajisje me ate te kërkuar nga AK-MSh.

2.2 Ankesa nga OE se "Must perform single and/or double" është e përfshirë ne faqen 3 te katalogut/broshurës:



Nuk qëndron pasi qe nga specifikat e DT është kërkuar qe pajisja te performoj Single dhe/ose (and/or) double, dmth OE duhet ta kuptoj se është kërkuar single edhe/ose double e jo vetëm single, ose me një spjegim me te mire po te ishte ofruar produkti Double atëherë do te ishte i pranueshëm por vetëm single jo. Dmth ne mendojm qe ankesa e OE nuk qendron.

6.2 Ankesa nga OE: Convenient servo-controlled stabilizer/CVT "Specifikimi 6.2 mbulohet nga specifikimi i UPS-së nuk qendron fare sepse UPS nuk e kryn funksionin e Convenient servo-controlled stabilizer/CVT. Veq nje spjegim per OE se qka ne kemi kerkuar.

CVT Voltage Stabilizer is used to control the voltage fluctuations and produce a desired and constant output which would enable machine to function without any interruption.

6.2 Convenient servo-controlled stabilizer/CVT, nuk keni dëshmuar me asnjë dokument (katallog, brochure apo specifikacion teknik) se pajisja e tille e përmbushë specifiken teknike te kërkuar. Deklarata juaj me vetëm "YES" nuk është e mjaftueshme.

8.2 Ankesa nga OE: List of equipment available to provide calibration and routine preventive maintenance support. according to the manufacturer's documentation in the service/technical manual nuk mjafton vetem me ***YES,*** mungon deklarata apo broshura e pajisjeve te kërkuara nga specifikimi I DT.

3.14 Apheresis sets (kits). Ju keni ofruar dokumentin "Technical Information Sheet 9400411 – P1R" si me poshtë:

FRESENIUS KABI

Technical Information Sheet
9400411 – P1R

Date: May, 2018



Reference: Product Name: Legal Manufacturer:	9400411 P1R Fresenius Kabi AG, 61346 Bad Homburg, Germany
Description:	P1R Plasma Treatment Set
Bag (filling) volumes:	1 * 500 ml Empty bag
Certification:	ISO 13485. CE marked: CE0123
Plastic Foil:	PVC-DEHP
Sterilisation:	ETO
Shelf life:	36 months
Packaging:	1 unit per package
Quantity per box:	5 units
Quantity per pallet:	120 units (24 boxes x 5 units)

The information on this sheet is subject to change, for actual information please contact your local representative/agent.

1/1

Nga skema e me sipërme shihet që mungon përshkrimi teknik sipas specifikave te DT-së, pra shihet qartë qe nuk ka ndonjë përmbajtje në dokument e cila dëshmon përmbushjen e specifikave nga DT.

LOT: 32

Komisioni i vlerësimi e ka rishqyrtuar edhe njëherë ofertën tuaj lidhur me lotin 32 dhe pas rishqyrtimin te dokumenteve i konfirmojmë këto mangësi ne specifika te ofertës tuaj:

	LOT 32	KEIS
	5. Vertical laminar air flow for IVF with length 180-200 cm with two heating and monitors	
	Quantity 1 pcs.	
	• It should be produced in a structure preventing vibration.	(Less Noise, Less Vibration)
	• In order to minimize noise and vibration, it must have a special EC dual fan system.	(ECM blower)
	• Device must have microprocessor control system.	(Esco Sentinel Gold Microprocessor Control System)
	• The manufacturer has ISO9001, ISO14001, device UL, NSF certificate, FDA In-Vitro Fertilization compliance certificate.	OE ka prezentuar vetem ISO9001, ISO13485 dhe EC Certfikatat, te tjerat nuk jane identifikuar ne dokumentacionin e dorzuar.
	6. Vertical laminar air flow for IVF with length 180-200 cm with heating and microscope housing place Quantity 2 pcs.	
	The device must have a homogeneous vertical flow laminar flow system (negative plenum) that creates positive and negative pressures.	No (Katallogu nuk lexohet)
	The power consumption of the device should be 600 Watt.	No (Katallogu nuk lexohet)
	The interior and sidewalls of the device must not contain welded parts, gaskets or screws that may cause contamination. Also, the side walls should be glazed.	No (Katallogu nuk lexohet)
	The interior of the device should be made of 304 stainless steel.	No (Katallogu nuk lexohet)
	The outer part of the device must be made of cold drawn steel plate resistant to corrosion.	No (Katallogu nuk lexohet)
	The working surface of the device must be one piece. Device's work table must be heated. There should be 1 stereo microscope housing on the device. The light source must be able to be attached to the device.	No (Katallogu nuk lexohet)
	The stand height of the device should be 650 mm.	No (Katallogu nuk lexohet)
	The outer dimensions of the device (WxDxH) should be approximately: 1900x700x2250 mm.	No (Katallogu nuk lexohet)
	The interior dimensions of the device (W x D x H) should be approximately: 1800x630x630 mm.	No (Katallogu nuk lexohet)
	The weight of the device should be approximately 150 kg.	No (Katallogu nuk lexohet)
	There should be a lower stand with the device as a standard.	No (Katallogu nuk lexohet)
	The device should be able to operate with 220 V, 50/60 Hz.	No (Katallogu nuk lexohet)
	Bottom stand suitable for the device should be given with the device.	No (Katallogu nuk lexohet)
	The device will be protected against fabrication and assembly faults for 3 years.	No (Katallogu nuk lexohet)

	7. Vertical laminar air flow for IVF with length 180-200 cm	
	Quantity 1 pcs.	
	The device must have a homogeneous vertical flow laminar flow system (negative plenum) that creates positive and negative pressures.	No (Katallogu nuk lexohet)
	The power consumption of the device should be 600 Watt.	No (Katallogu nuk lexohet)
	The interior and sidewalls of the device must not contain welded parts, gaskets or screws that may cause contamination. Also, the side walls should be glazed.	No (Katallogu nuk lexohet)
	The interior of the device should be made of 304 stainless steel.	No (Katallogu nuk lexohet)
	The outer part of the device must be made of cold drawn steel plate resistant to corrosion.	No (Katallogu nuk lexohet)
	The working surface of the device must be one piece. Device's work table must be heated. The table should be flat.	No (Katallogu nuk lexohet)
	The stand height of the device should be 650 mm.	No (Katallogu nuk lexohet)
	The outer dimensions of the device (W x D x H) should be approximately: 1900x700x2250 mm.	No (Katallogu nuk lexohet)
	The interior dimensions of the device (W x D x H) should be approximately: 1800x630x630 mm.	No (Katallogu nuk lexohet)
	There should be a lower stand with the device as a standard.	No (Katallogu nuk lexohet)
	The device should be able to operate with 220 V, 50/60 Hz.	No (Katallogu nuk lexohet)
	Bottom stand suitable for the device should be given with the device.	No (Katallogu nuk lexohet)
	The device will be protected against fabrication and assembly faults for 3 years.	No (Katallogu nuk lexohet)

Pas ankesës tuaj ne i kemi verifikuar edhe një here katalogun nga web-site i prodhuesit:

Ne kete web-site nuk egziston ndonje katalog per artikujt 6 dhe 7 te lotit 32.

Ne web-site kemi evidentuar nje katalog ne gjuhen turke, mirepo as ne katalogut ne gjuhen Turke nuk egzistojne keta dy artikuj.

LOT: 33

Perkunder tejkalimit te buxhetit, komisioni vleresues ka bere edhe nje here vlersimin e produktit te ofruar dhe konsiderojme se:

Ju keni ofertuar një broshure tejet bazike për Softuerin, e cila nuk përmban as 20% te specifikave teknike te kërkuara nga dosja e tenderit, andaj konsiderojmë se produkti i ofruar nuk i përmbush kërkesat minimale te përcaktuar tek Specifikimet Tekniket te DT-se.
Mungojnë disa nga specifikat e me poshtme ne dokumentacionin e ofruar:

SOFTWARE

User interface and chip reader software have been developed using C #, .Net, Java and C ++ programming languages. User interfaces have been developed in a standard to be used in the web platform. Critical patient data are kept encrypted in the database. The language of the developed management software must be Albanian or English language. Additional languages can be added on request. It is used by connecting to the main server via a web browser. In the software; It includes Patient Card, Personnel Card, Laboratory (Andrology, Embryology, OPU, ICSI, Control, Transfer, Cryo, Thaw), Users and Authorization modules and detailed reporting. All definitions (patient, staff) and authorizations can be made via software. In addition, all patient movements (procedure type, operating personnel, processing device, hour, time, etc.) and personnel movements can be reported.

SYSTEM TECHNICAL REQUIREMENTS

- In the relevant location where the system will operate, there must be a PC server computer with a minimum of Windows 2016 Server operating system installed or better version of Windows. In addition, this server must have a database program with a minimum Microsoft SQL 2017 version and an IIS web platform or better versions.

Nuk ka informacion ne dokumentacionin e ofruar për pjesën e Moduleve siç janë kërkuar ne DT (1- APPOINTMENT MODULES, 2- PATIENT ACCEPTANCE, etj)

LOT: 34

Pas shqyrtimit te ankese se OE komisioni vlerësues konstaton se produktet e ofruara nuk i përmbushin specifikat teknike te kërkuara si ne vijim:

Produktet e ofruara janë të selektuara vetëm modelet por nuk janë te nënvizuara specifikat teknike, mungon linku i katalogut, nuk plotësohet Kërkesa 2 nga pika 9.1 e FTD-së. Megjithatë vlerësimi është bërë në bazë të gjetjeve nëpër katalog, gjeni disa nga specifikat qe nuk janë plotësuar:

1. ***Tek Artikulli: 11. CO2 Incubator (volume 160-170 liters) mungon specifikat:***
   a. ***IRO2 Sensor 01-21%***
   b. ***Direct heated water humidity reservoir 3l***
   c. ***Partitioned internal three glass gas tight doors***
   d. ***With heat sterilization cycle on 180oC high temperature in period of 12 h with automatic lock during sterilization***
2. ***Tek Artikulli: 14. Cryo shipper Dewar for storage eshte kërkuara Static holding min 21 days, ju keni ofruar Vapor Static holdover time (Day) 14, devijimi është më i madh se 40% e vlerës se kërkuar.***
3. ***Tek Artikulli: 17. Set of automatic pipettes, mungojnë Increments / Accuracy / Precision.***

Zyrtari Përgjegjës i prokurimit bazuar ne ato që u thane më lart ka vendosur si ne dispozitiv.

Këshilla juridike:

Kundër këtij vendimi, ankuesi mund të paraqesë ankesë pranë OSHP-së. **Ankesa prane OSHP-se duhet te dorezohet brenda 10 (dhjete) diteve pas pranimit te ketij vendimi.**

Vendimi i dorëzohet:

1. KEIS Pharmaceutical Sh.P.K
2. Operatoreve ekonomik te interesuar

Berat Marmullaku

Zyrtar përgjegjës i Prokurimit